

**GOVERNMENT OF INDIA
DEPARTMENT OF SPACE
INDIAN INSTITUTE OF REMOTE SENSING
IIRS
DEHRADUN
PURCHASE & STORES
INVITATION TO TENDER**

Ph No: 0135 - 2524317, 4318

Fax 0135 - 2748041

Email: pns@iirs.gov.in

Date :23/07/2014

M/s

000000

Our Ref No : GIER 2014-000106-01

Tender Due: 15:00 Hrs ISTon 14/08/2014

Dear Sirs,

Please submit your sealed quotation , in the Tender Form enclosed here along with the descriptive catalogues / pamphlets / literature ,superscribed with Our Ref.No. and Due Date for the supply of the following items as per the terms & conditions mentioned in Annexure(Form No:)

S.No.	Description of Items with Specifications	Unit	Quantity
1	Variable Volume Electronic Single Channel Micropipettes, Volume Range: 100-5000µl with 10µl increment, Specifications: 1- Accuracy about +- 0.5-1.6% over the range. 2- calibration & Accuracy as per ISO8655	NOS.	1
2	Electronic Pipette Filling Devices, Collet & Plunger 220 V	NOS.	1
3	Fixed Volume single channel Micropipettes Specification: 1- Accuracy about +- 0.3 over the range. 2- Fully Autoclavable, calibration & Accuracy as per ISO 8655, Volume 50µl	NOS.	1
4	Fixed Volume single channel Micropipettes Specification: 1- Accuracy about +- 0.3 over the range. 2- Fully Autoclavable, calibration & Accuracy as per ISO 8655, Volume 100µl	NOS.	1
5	Fixed Volume single channel Micropipettes Specification: 1- Accuracy about +- 0.3 over the range. 2- Fully Autoclavable, calibration & Accuracy as per ISO 8655, Volume 200µl	NOS.	1
6	Fixed Volume single channel Micropipettes Specification: 1- Accuracy about +- 0.3 over the range. 2- Fully Autoclavable, calibration & Accuracy as per ISO 8655, Volume 250µl	NOS.	1
7	Fixed Volume single channel Micropipettes Specification: 1- Accuracy about +- 0.3 over the range. 2- Fully Autoclavable, calibration & Accuracy as per ISO 8655, Volume 500µl	NOS.	1
8	Fixed Volume single channel Micropipettes Specification: 1- Accuracy about +- 0.3 over the range. 2- Fully Autoclavable, calibration & Accuracy as per ISO 8655, Volume 1000µl	NOS.	1
9	Fixed Volume single channel Micropipettes Specification: 1- Accuracy about +- 0.3 over the range. 2- Fully Autoclavable, calibration & Accuracy as per ISO 8655, Volume 2000µl	NOS.	1
10	Fixed Volume single channel Micropipettes Specification: 1- Accuracy about +- 0.3 over the range. 2- Fully Autoclavable, calibration & Accuracy as per ISO 8655, Volume 5000µl	NOS.	1
11	Fixed Volume single channel Micropipettes Specification: 1- Accuracy about +- 0.3 over the range. 2- Fully Autoclavable, calibration & Accuracy as per ISO 8655, Volume 10000µl	NOS.	1
12	Variable Volume single channel Micropipettes, Specifications 1- Accuracy about +- 0.5-1.6% over the range. 2- Fully Autoclavable, Calibration & Accuracy as per ISO 8655, Volume range 100-1000µl with 10 µl increment	NOS.	1
13	Variable Volume single channel Micropipettes, Specifications 1- Accuracy about +- 0.5-1.6% over the range. 2- Fully Autoclavable, Calibration & Accuracy as per ISO 8655, Volume range 500-5000µl with 100 µl increment	NOS.	1
14	Variable Volume single channel Micropipettes, Specifications 1- Accuracy about +- 0.5-1.6% over	NOS.	1

01c

**GOVERNMENT OF INDIA
DEPARTMENT OF SPACE
INDIAN INSTITUTE OF REMOTE SENSING
IIRS**

Ph No: 0135 - 2524317, 4318

Fax 0135 - 2748041

Email: pns@iirs.gov.in

DEHRADUN

PURCHASE & STORES

Date :23/07/2014

the range. 2- Fully Autoclavable, Calibration & Accuracy as per ISO 8655, Volume range 1000-10000µl with 100µl increment

15	Variable Volume single channel Micropipettes with 1µl increment, Specifications: 1- Accuracy about +- 0.7-2.5% over the range. 2- Fully Autoclavable, Calibration & Accuracy as per ISO 8655, Volume range 5-50µl	NOS.	1
16	Variable Volume single channel Micropipettes with 1µl increment, Specifications: 1- Accuracy about +- 0.7-2.5% over the range. 2- Fully Autoclavable, Calibration & Accuracy as per ISO 8655, Volume range 10-100µl	NOS.	1
17	Variable Volume Multichannel (8 Channel) Micropipettes with 1µl increment, Specifications 1- Accuracy about +- 0.7-3% over the range. 2- Fully Autoclavable, Calibration & Accuracy as per ISO 8655, Volume range 10-100µl	NOS.	1
18	Stands for Micropipette, 1- Micropipette Stand for 3 single channel pipettes	NOS.	2
19	Stands for Micropipette, 1- Micropipette Stand for 6 single channel pipettes	NOS.	2
20	Stands for Micropipette, 1- Micropipette Stand for 8 Multi channel pipettes	NOS.	1

DELIVERY AT: IIRS

MODE OF DESPATCH DOOR DLVRY

DUTY EXEMPTIONS

SPECIAL INSTRUCTIONS NIL

SPECIFIC TERMS


V.V. NARAYANAN KUTTY
PURS. & STORES OFFICER

For and on behalf of the President of India
The Purchaser

Terms & Conditions:

1. Material should be delivered & installed at IIRS
2. Payment will be made within 30 days from date of receipt of supply and acceptance of the material for orders value upto Rs.2.0 lakhs.
3. We cannot furnish Form C/D. Please indicate the applicable percentage of Trade tax / VAT in your quotation, if applicable. Otherwise the quoted rate will be considered as inclusive of all taxes.
4. Clearly mention the Make / Brand of the item in your quotation. Please enclosed the Authorization Certificate from the principal of the quoted Make / Model along with the quotation.
5. Also clearly mention the exact delivery period and validity of your offer shall be min 60 days.