

**GOVERNMENT OF INDIA
DEPARTMENT OF SPACE
INDIAN INSTITUTE OF REMOTE SENSING
DE-IIRS
PURCHASE & STORES**

Ph No: 0135 - 2524317, 4318
Fax 0135 - 2748041
Email: pns@iirs.gov.in

Date :08/06/2016

INVITATION TO TENDER

M/s

000000

Our Ref No : GIER 2016-000072-01

Tender Due: 15:00 Hrs ISTon 30/06/2016

Dear Sirs,

Please submit your sealed quotation , in the Tender Form enclosed here along with the descriptive catalogues / pamphlets /literature ,superscribed with Our Ref.No. and Due Date for the supply of the following items as per the terms & conditions mentioned in Annexure(Form No:)

S.No.	Description of Items with Specifications	Unit	Quantity
1	Pressure base Water Level Recorder, Specifications: Suitable for measuring ground water & surface water level, 2- It should have a pressure sensor, temperature detector, data logger & minimum 5 Years battery life (Detailed Specification enclosed)	NOS.	1


DELIVERY AT: IIRS

MODE OF DESPATCH DOOR DLVRY

DUTY EXEMPTIONS

SPECIAL INSTRUCTIONS NIL

SPECIFIC TERMS


V.V. NARAYANAN KUTTY
PURS. & STORES OFFICER
For and on behalf of the President of India
The Purchaser

Note- Fax/E-mail quotation will not be considered.

Terms & Conditions:

1. Material should be delivered & installed at IIRS
2. Payment will be made within 30 days from date of receipt of supply and acceptance of the material for orders value upto Rs.2.0 lakhs.
3. We cannot furnish Form C/D. Please indicate the applicable percentage of Trade tax / VAT in your quotation, if applicable. Otherwise the quoted rate will be considered as inclusive of all taxes.
4. Clearly mention the Make / Brand of the item in your quotation. Please enclosed the Authorization Certificate from the principal of the quoted Make / Model along with the quotation.
5. Also clearly mention the exact delivery period and validity of your offer shall be min 60 days.

भारत सरकार
अन्तरिक्ष विभाग
भारतीय सुदूर संवेदन संस्थान
देहरादून
क्रय एवं भंडार अनुभाग
निविदा आमंत्रण

टेलीफोन: 0135-2524317, 4318
फैक्स: 0135-2748041
ईमेल: prs@iirs.gov.in

संदर्भ सं. जी.आई.ई.आर. 2016-000072-01

दिनांक: 08/06/2016

निविदा जमा करने की अंतिम तिथि: 30.06.2016 (1500 बजे)

सेवा में,

महोदय,

1. इस संस्थान द्वारा निम्नलिखित सामग्री का क्रय किया जाना है। आपसे आग्रह है कि उक्त सामग्री की पूर्ति हेतु संलग्न निविदा प्रपत्र में अपनी मुहरबंद बोली जमा करने का कष्ट करें। उक्त बोली के साथ यथावश्यक कैटलॉग/पैम्फलेट/लिटरेचर आदि भी जमा करें। मुहरबंद लिफाफे पर ऊपर की ओर संदर्भ संख्या तथा अंतिम तिथि अंकित होनी चाहिए। सामग्री की पूर्ति से संबंधित निबंधन एवं शर्तें संलग्नक (प्रपत्र सं.....) में वर्णित हैं।

क्रम सं.	सामग्री का विवरण तथा विनिर्देश	इकाई	संख्या
1.	Pressure base Water Level Recorder, Specifications: Suitable for measuring ground water & surface water level, 2- It should have a pressure sensor, temperature detector, data logger & minimum 5 Years battery life (Detailed Specification enclosed)	Nos.	01


डिलिवरी का स्थान : भारतीय सुदूर संवेदन संस्थान

प्रेषण की विधि : द्वार पर

शुल्क छूट :

विशेष निर्देश : कुछ नहीं

विशिष्ट शर्तें :


वी.वी. नारायणन कुट्टी
क्रय एवं भंडार अधिकारी
कृते भारत के राष्ट्रपति,
क्रेता

नोट: फैक्स/ईमेल द्वारा भेजी गई निवेदित दर/कोटेशन पर विचार नहीं किया जाएगा।

निबंधन एवं शर्तें

- सामग्री की डिलिवरी तथा संस्थापन भारतीय सुदूर संवेदन संस्थान में किया जाना चाहिए।
- रु0 2 लाख मूल्य तक के क्रय आदेशों का भुगतान सामग्री की प्राप्ति एवं स्वीकरण तिथि से एक माहके भीतर किया जाएगा।
- प्रपत्र/फॉर्म सी/डी इस संस्थान द्वारा जारी नहीं किया जा सकता। यथा लागू, अपनी निवेदित दर/कोटेशन में व्यापार कर/वैट के प्रतिशत का उल्लेख अवश्य करें, अन्यथा, निवेदित दर/कोटेशन में समस्त करों को सम्मिलित माना जाएगा।
- अपनी निवेदित दर/कोटेशन में सामग्री की मेक/निर्माता/ब्रांड का उल्लेख स्पष्टया करें। कृपया अपनी निवेदित दर/कोटेशन के साथ उद्धृत मेक/निर्माता/ब्रांड के प्रमुख से प्राधिकरण प्रमाण-पत्र भी संलग्न करें।
- सटीक डिलिवरी अवधि तथा आपके प्रस्ताव की न्यूनतम वैधता 60 दिन होने का भी उल्लेख स्पष्टतया करें।

Detailed Tech. Details / Catalogue / Data Sheet Required : Yes

Technical specification:

Level Sensor:	Highly durable Pressure Sensor
Ranges/ Accuracy:	5M or 10M; Accuracy: 0.1% FS
Units of Measure:	cm, m, ft & °C
Normalization:	Automatic Temp Compensation between 0°C to 40°C
Temperature Sensor:	RTD
Range/Resolution/Accuracy:	60°C /0.1°C/ $\pm 0.1^{\circ}\text{C}$
Battery life:	Minimum 5 Years (based on 1 reading/minute)
Operating Temperature:	- 10°C to 60°C
Clock Accuracy:	± 1 minute/year (- 20°C to 80°C)
Memory:	>30,000 sets of readings
Communication:	Through USB & wirelessly to Smart phones/Tablets with user friendly software
Size/ Weight:	Suitable for installation even in 1" dia well
Measurement Rates	0.5 sec to 99 hours
Barometric Compensation:	Through Barometric Pressure sensor